

Manual Supplement

Manual Title:	32X Calibration	Supplement Issue:	2
Part Number:	1631636	Issue Date:	4/09
Print Date:	October 2001	Page Count:	1
Revision/Date:	1, 12/03		

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1

On page 9, Table 3, replace the **DC Volts** section, with the following:

Functional Test	5520A Output	Nominal	Low Limit	High Limit
DC Volts	-350 V	-350.0 V	-354.0	-346.0
	35 V	35.0 V	34.1	35.9
	350 V	350.0 V	346.0	354.0
	600 V	600 V	589	611

Change #2 w2030084

On page 2, replace the TUV symbol with the following:

